**Unit no. HVS-06**

IKM Testing AS

# WATER ANALYSER /-SENSOR

## *Dimensions*

Length(mm).....: 180 Width(mm).....: 270 Height(mm).....: 100
Weight (kg).....: 2

## *Construction year*

Built.....: 2007

## *Type*

Item type.....: Measure water content in oil
Fabricator.....: Pall
Model.....: WS-04
WSPC06
Serial no.....: Pall WS-04: Ser. No: 2039
Pall WSPC06 Ser. No: 4432

## *Power supply*

Battery.....: 9V Battery

## *Content*

Content.....: 1 ea. Pall WS-04
1 ea. Pall WSPC06
1 ea. Extra battery
1 ea. 3/8” fitting

## *Flow / Pressure / Volume*

Measure Range.....: 0-100% saturation / ppm

## *Additional information*

Equipment / Functions.....: Measure water content in oil. (Petroleum and synthetic fluids and fuels)

To read out results in ppm you need to program the unit with an equation for the specific oil. Must be specified in Order