Unit no. UM-16

IKM Testing AS

ACOUSTIC LEAK DETECTOR

Dimensions

Length(mm).....: 250 Width(mm).....: 100 Height(mm).....: 250
Weight (kg).....: 2

Construction year

Built.....: 2007

Area of use / certification

Standard.....: CE

Type

Item type.....: Leak detection in pipes with compressed air or water, ball-bearing, motor defect
Fabricator.....: Sewerin
Model.....: Stethophon 04

Power supply

Battery.....: 9 V, 6LR61 (Approx. 14 hours operating time)

Content

Content.....: 1 ea. Sewerin Stethophon P
1 ea. Sony headset

Flow / Pressure / Volume

Minimum temp. (GrC).....: -10
Max temp. (GrC).....: 50
Sound amplify.....: 13000 times (82 dB)